



Script[®] One-Step qRT-PCR Kit

Cat. No 95057-050 Size: 50 x 50- μ L reactions
95057-200 200 x 50- μ L reactions

Store at -25°C to -15°C

Description

The qScript One-Step qRT-PCR Kit is a convenient and highly sensitive solution for reverse transcription quantitative PCR (RT-qPCR) of RNA templates using hybridization probe detection chemistries such as TaqMan[®] 5'-hydrolysis probes or molecular beacons on real-time quantitative PCR systems that do not require an internal reference dye. cDNA synthesis and PCR amplification are carried out in the same tube without opening between procedures. The system has been optimized to deliver maximum RT-PCR efficiency, sensitivity, and specificity, enabling unbiased co-amplification of low copy transcripts in the presence of higher copy reference genes. The proprietary reaction buffer has been specifically formulated to maximize activities of both reverse transcriptase and Taq DNA polymerase while minimizing the potential for primer-dimer and other non-specific PCR artifacts. Highly specific amplification is crucial to successful qRT-PCR as non-specific product(s) can compete for amplification of the target sequence and impair PCR efficiency. A key component of this kit is AccuStart Taq DNA polymerase, which contains monoclonal antibodies that bind to the polymerase and keep it inactive prior to the initial PCR denaturation step. Upon heat activation at 95°C, the antibodies denature irreversibly, releasing fully active, unmodified Taq DNA polymerase.

Instrument Compatibility

Different real-time PCR systems employ different strategies for the normalization of fluorescent signals and correction of well-to-well optical variations. It is critical to match the appropriate qPCR reagent to your specific instrument. The qScript One-Step qRT-PCR Kit does not contain an internal reference dye. Please consult the following table, or visit our web site at www.quantabio.com to find an optimized kit for your instrument platform(s).

Reagent	Cat Nos	Compatible Real-Time PCR Systems
qScript One-Step qRT-PCR Kit, ROX	95058-050 95058-200	Applied Biosystems 7000, 7300, 7700, 7900, 7900HT, 7900HT Fast, StepOne [™] , StepOnePlus [™]
qScript One-Step qRT-PCR Kit, Low ROX	95059-050 95059-200	Applied Biosystems 7500, 7500 Fast, ViiA [™] 7 Stratagene MX4000 [™] , MX3005P [™] , MX3000P [™]
qScript One-Step qRT-PCR Kit	95057-050 95057-200	Bio-Rad CFX96 [™] , CFX384 [™] , iCycler iQ [®] , iQ [™] 5, MyiQ [™] Opticon [™] , MiniOpticon [™] , Chromo4 [™] Cepheid Smart Cycler [®] ; Qiagen/Corbett Rotor-Gene [®] Eppendorf Mastercycler [®] ep realplex Roche Applied Science LightCycler [®] 480

Components

Reagent

Reagent	Description
qScript One-Step Reverse Transcriptase	Optimized 50X formulation of recombinant MMLV reverse transcriptase for one-step RT-PCR.
One-Step Master Mix (2X)	2X reaction buffer containing dNTPs, magnesium chloride, AccuStart Taq DNA polymerase, and stabilizers
Nuclease-free water	

Storage and Stability

Store components in a constant temperature freezer at -25°C to -15°C upon receipt.

Repeated freezing and thawing of the reaction mix is not recommended.

For lot specific expiry date, refer to package label, Certificate of Analysis or Product Specification Form.

Guidelines for One-Step qRT-PCR

- Thaw all components, except qScript One-Step RT, at room temperature. Mix vigorously, and then centrifuge to collect contents to the bottom of the tube before using. Place all components on ice after thawing.
- To maximize specificity, reactions should be assembled on ice. AccuStart Taq DNA polymerase is inactive prior to high temperature activation; however, qScript One-Step reverse transcriptase is active at lower temperatures. First-strand synthesis can be carried out between 42°C and 52°C. Optimal results are generally obtained with a 10-minute incubation at 48 – 50°C. We recommend a 5 minute incubation at 95°C to fully inactivate the RT prior to PCR cycling.
- Preparation of a reaction cocktail is recommended to reduce pipetting errors and maximize assay precision. Assemble the reaction cocktail with all required components except RNA template and dispense equal aliquots into each reaction tube. Add RNA to each reaction as the final step. Addition of sample as 5 to 10-µL volumes will improve assay precision.
- Suggested input quantities of template are: 1 pg to 1 µg total RNA; 10 fg to 100 ng poly A(+) RNA; 10 to 1x10⁸ copies viral RNA.
- After sealing each reaction, vortex gently to mix contents. Centrifuge briefly to collect components at the bottom of the reaction tube.

Reaction Assembly

Component	Volume for 50-µL rxn.	Final Concentration
One-Step Master Mix (2X)	25 µL	1X
Forward primer	Variable	400 – 900 nM
Reverse primer	Variable	400 – 900 nM
Probe	Variable	50-200 nM
Nuclease-free water	Variable	
RNA template	5 – 10 µL	Variable
qScript One-Step RT	<u>1 µL</u>	1X
Final Volume (µL)	50 µL	

Note: For smaller reaction volumes (i.e. 25-µL reactions), scale all components proportionally.

Reaction Protocol

Incubate complete reaction mix in a real-time thermal detection system as follows:

cDNA Synthesis	48 – 50°C, 10 min
Initial denaturation	95°C, 5 min
PCR cycling (30 - 45 cycles)	95°C, 10 to 15s
	55 – 60°C, 30 to 60s (data collection step)

Quality Control

Kit components are free of contaminating DNase and RNase. The qScript One-Step qRT-PCR Kit is functionally tested in RT-qPCR. Kinetic analysis must demonstrate linear resolution over six orders of dynamic range ($r^2 > 0.995$) and a PCR efficiency > 90%

Limited Label Licenses

Use of this product signifies the agreement of any purchaser or user of the product to the following terms:

1. The product may be used solely in accordance with the protocols provided with the product and this manual and for use with components contained in the kit only. QIAGEN Beverly, Inc. grants no license under any of its intellectual property to use or incorporate the enclosed components of this kit with any components not included within this kit except as described in the protocols provided with the product, this manual, and additional protocols available at www.quantabio.com. Some of these additional protocols have been provided by Quantabio product users. These protocols have not been thoroughly tested or optimized by QIAGEN Beverly, Inc.. QIAGEN Beverly, Inc. neither guarantees them nor warrants that they do not infringe the rights of third-parties.
2. Other than expressly stated licenses, QIAGEN Beverly, Inc. makes no warranty that this kit and/or its use(s) do not infringe the rights of third-parties.
3. This kit and its components are licensed for one-time use and may not be reused, refurbished, or resold.
4. QIAGEN Beverly, Inc. specifically disclaims any other licenses, expressed or implied other than those expressly stated.
5. The purchaser and user of the kit agree not to take or permit anyone else to take any steps that could lead to or facilitate any acts prohibited above. QIAGEN Beverly, Inc. may enforce the prohibitions of this Limited License Agreement in any Court, and shall recover all its investigative and Court costs, including attorney fees, in any action to enforce this Limited License Agreement or any of its intellectual property rights relating to the kit and/or its components.

©2018 QIAGEN Beverly Inc. 100 Cummings Center Suite 407J Beverly, MA 01915

Quantabio brand products are manufactured by QIAGEN, Beverly Inc.

Intended for molecular biology applications. This product is not intended for the diagnosis, prevention or treatment of a disease.

qScript is a registered trademark of QIAGEN Beverly, Inc.

TaqMan is a registered trademark of Roche Molecular Systems, Inc. LightCycler is a registered Trademark of Roche. Applied Biosystems, StepOne, StepOnePlus, ViiA, and ROX are trademarks Life Technologies Corporation. Stratagene, MX3000P, MX3005P and MX4000 are trademarks of Agilent Technologies, Inc. Mastercycler is a trademark of Eppendorf. Rotor-Gene is a registered trademark of Qiagen GmbH. SmartCycler is a trademark of Cepheid. CFX96, CFX384, iCycler iQ, iQ5, MyiQ, Opticon, MiniOpticon and Chromo4 are trademarks of Bio-Rad Laboratories.