

## PerfeCTa<sup>®</sup> MultiPlex qPCR SuperMix

Cat. No	95063-050	Size:	50 x 50- $\mu$ L reactions
	95063-200		200 x 50- $\mu$ L reactions
	95063-01K		1000 x 50- $\mu$ L reactions

Store at -25°C to -15°C  
protected from light

### Description

PerfeCTa MultiPlex qPCR SuperMix is a 2X concentrated, ready-to-use reaction cocktail for real-time quantitative PCR (qPCR) that contains all components, except primers, probes and templates. The system transcends multiplex limitations of conventional PCR master mixes, enabling unbiased amplification of up to five target sequences in a single tube. Suppression of low copy amplicons by high copy reference targets in the amplification is a common problem in multiplex PCR. This can skew, or mask the apparent representation and quantification of low copy target sequences. PerfeCTa MultiPlex qPCR SuperMix delivers dynamic range and sensitivity to multiplexed qPCR that is comparable to that for single-plex qPCR probe assays without the need for limiting or variable primer concentrations.

A key component of this supermix is AccuStart<sup>™</sup> Taq DNA polymerase, which contains monoclonal antibodies that bind to the polymerase and keep it inactive prior to the initial PCR denaturation step. Upon heat activation (2 minutes at 95°C), the antibodies denature irreversibly, releasing fully active, unmodified Taq DNA polymerase. This enables specific and efficient primer extension with the convenience of room temperature reaction assembly.

### Instrument Compatibility

Different real-time PCR systems employ different strategies for the normalization of fluorescent signals and correction of well-to-well optical variations. It is critical to match the appropriate qPCR reagent to your specific instrument. PerfeCTa MultiPlex qPCR SuperMix does not contain an internal reference dye to allow greater flexibility in your choice of reporter fluorophores and instrument platforms. Concentrated solutions of ROX<sup>™</sup> Reference Dye or Low ROX Reference Dye are provided separately. In general, instruments that utilize variable excitation wavelengths that are tuned to each respective dye detection channel provide superior sensitivities and dynamic ranges for multiplex probe applications. Your choice of probe reporter dyes and any optional internal reference dye must be matched to the excitation and emission optics of your particular instrument. Please consult the user manual for your real-time PCR system.

### Components

PerfeCTa Multiplex qPCR Supermix	2X reaction buffer containing optimized concentrations of MgCl <sub>2</sub> , dNTPs (dATP, dCTP, dGTP, dTTP), AccuStart Taq DNA Polymerase, and stabilizers.
ROX Reference Dye (50X)	50X concentrated ROX solution optimized for Applied Biosystems 7000, 7300, 7700, 7900, 7900HT, 7900HT Fast, StepOne <sup>™</sup> , or StepOnePlus <sup>™</sup>
Low ROX Reference Dye (50X)	50X concentrated ROX solution optimized for systems using 580 nm to 585 nm excitation wavelength for the ROX dye channel: Applied Biosystems 7500, 7500 Fast, ViA <sup>™</sup> 7, or Stratagene MX4000 <sup>™</sup> , MX3005P <sup>™</sup> , MX3000P <sup>™</sup>

### Storage and Stability

Store components in a constant temperature freezer at -25°C to -15°C protected from light upon receipt.

Repeated freezing and thawing of the Supermix is not recommended.

For lot specific expiry date, refer to package label, Certificate of Analysis or Product Specification Form.

### Guidelines for Multiplex qPCR:

- The design of highly specific primers and probes is a critical and challenging aspect of successful multiplex qPCR. Each primer and probe should have similar thermodynamic properties to support efficient PCR amplification using a common temperature cycling program for all amplicons. The use of computer aided primer design programs is encouraged in order to minimize the potential for internal secondary structure and complementation at 3'-ends within each primer, primer pairs, and primer/probe combinations
- Amplicon size should be consistent for each target sequence and limited to approximately 65 - 100 bp.

### Guidelines for qPCR continued:

- Limiting primer concentration for high copy genes is acceptable, but not required. A final concentration of 300 nM each primer and 100 to 250 nM probe is effective for most applications. Each probe for a multiplex assay should be labeled using dyes with minimal spectral overlap and non-fluorescent quencher compounds. Matching dyes with discrete fluorescent excitation and emission optima improves the accuracy of the multicomponenting, or dye deconvolution algorithms employed by the real-time PCR analysis software.
- Preparation of a reaction cocktail is recommended to reduce pipetting errors and maximize assay precision. Assemble the reaction cocktail with all required components except sample template (genomic DNA or cDNA) and dispense equal aliquots into each reaction tube. Add the DNA template to each reaction as the final step. Addition of samples as 5 to 10- $\mu$ L volumes will improve assay precision.
- Suggested input quantities of template are: cDNA corresponding to 10 pg to 1  $\mu$ g of total RNA; 100 pg to 1  $\mu$ g genomic DNA
- After sealing each reaction, vortex gently to mix contents. Centrifuge briefly to collect components at the bottom of the reaction tube.

### Reaction Assembly

Component	Volume for 50- $\mu$ L rxn.	Final Concentration
PerfeCTa Multiplex qPCR SuperMix (2X)	25 $\mu$ L	1x
Forward primers	variable	100 – 500 nM
Reverse primers	variable	100 – 500 nM
Probes	variable	100 – 250 nM
ROX or Low ROX Reference Dye (50X)	1 $\mu$ L	<i>optional</i>
Nuclease-free water	variable	
Template(s)	<u>5 – 10 <math>\mu</math>L</u>	variable
Final Volume ( $\mu$ L)	50 $\mu$ L	

### Reaction Protocol

Incubate complete reaction mix in a real-time thermal detection system as follows:

Initial denaturation:	95°C, 2 to 3 min
PCR cycling (30-45 cycles):	95°C, 10 to 15 s 55 – 65°C, 60s (collect and analyze data)

Full activation of AccuStart Taq DNA polymerase occurs within 30 seconds at 95°C. Initial denaturation times greater than 3 minutes are usually not required. However, amplification of gDNA targets may benefit from a prolonged initial denaturation step (5-10 min) to fully denature and fragment the template. This minimizes the potential for renaturation of long fragments and/or repetitive sequence regions that can impair replication of the target sequence by the PCR process.

### Quality Control

Kit components are free of contaminating DNase and RNase. PerfeCTa Multiplex qPCR SuperMix is functionally tested in a four-plex TaqMan qPCR using variable concentrations of one target sequence from 100 to 1 x 10<sup>7</sup> copies and 1 x 10<sup>8</sup> copies each of three other target sequences. Kinetic analysis must demonstrate linear resolution over six orders of dynamic range ( $r^2 > 0.995$ ) and a PCR efficiency > 90%.

### Limited Label Licenses

Use of this product signifies the agreement of any purchaser or user of the product to the following terms:

1. The product may be used solely in accordance with the protocols provided with the product and this manual and for use with components contained in the kit only. QIAGEN Beverly, Inc. grants no license under any of its intellectual property to use or incorporate the enclosed components of this kit with any components not included within this kit except as described in the protocols provided with the product, this manual, and additional protocols available at [www.quantabio.com](http://www.quantabio.com). Some of these additional protocols have been provided by Quantabio product users. These protocols have not been thoroughly tested or optimized by QIAGEN Beverly, Inc.. QIAGEN Beverly, Inc. neither guarantees them nor warrants that they do not infringe the rights of third-parties.
2. Other than expressly stated licenses, QIAGEN Beverly, Inc. makes no warranty that this kit and/or its use(s) do not infringe the rights of third-parties.
3. This kit and its components are licensed for one-time use and may not be reused, refurbished, or resold.
4. QIAGEN Beverly, Inc. specifically disclaims any other licenses, expressed or implied other than those expressly stated.
5. The purchaser and user of the kit agree not to take or permit anyone else to take any steps that could lead to or facilitate any acts prohibited above. QIAGEN Beverly, Inc. may enforce the prohibitions of this Limited License Agreement in any Court, and shall recover all its investigative and Court costs, including attorney fees, in any action to enforce this Limited License Agreement or any of its intellectual property rights relating to the kit and/or its components.

©2018 QIAGEN Beverly Inc. 100 Cummings Center Suite 407J Beverly, MA 01915

Quantabio brand products are manufactured by QIAGEN, Beverly Inc.

Intended for molecular biology applications. This product is not intended for the diagnosis, prevention or treatment of a disease.

PerfeCTa and AccuStart are trademarks of QIAGEN Beverly, Inc. TaqMan is a registered trademark of Roche Molecular Systems, Inc. LightCycler is a registered Trademark of Roche. Applied Biosystems, StepOne, StepOnePlus, ViiA, and ROX are trademarks Life Technologies Corporation. Stratagene, MX3000P, MX3005P and MX4000 are trademarks of Agilent Technologies, Inc.