



PerfeCTa® qPCR FastMix®, UNG, Low ROX™

Cat. No.	95078-250 95078-012 95078-05K	Size: 250 x 20-µL reactions (2 x 1.25 mL) 1250 x 20-µL reactions (10 x 1.25 mL) 5000 x 20-µL reactions (1 x 50 mL)	Store at -20°C protected from light
----------	-------------------------------------	---	-------------------------------------

Description

PerfeCTa qPCR FastMix, UNG, Low ROX is a 2X concentrated, ready-to-use reaction cocktail that contains all components, except primers, probe(s), and template for real-time quantitative PCR on Applied Biosystems 7500, 7500 Fast, ViiA™ 7 or Stratagene MX series of real-time PCR systems. The proprietary buffer and stabilizers have been specifically optimized to deliver maximum PCR efficiency, sensitivity, and robust fluorescent signal with TaqMan® or TaqMan MGB probe chemistry when using rapid PCR cycle times and reduced reaction volumes. This affords greater reagent economy and laboratory throughput on conventional or rapid ramp rate qPCR systems. The enhanced specificity of this FastMix suppresses cross-reactivity between homologous sequences, improving detection and discrimination in SNP applications. A key component of this FastMix is AccuFast™ Taq DNA polymerase. This hot-start Taq contains a proprietary mixture of monoclonal antibodies that bind to the polymerase and keep it inactive prior to the initial PCR denaturation step (> 48 hours at room temperature). Similar to our AccuStart™ Taq DNA polymerase, these antibodies are irreversibly inactivated during the initial PCR denaturation step. However, unlike other antibody hot-start polymerases, activation of AccuFast Taq is instantaneous at 95°C. Rapid recovery of fully active, unmodified Taq DNA polymerase is critical for efficient extension kinetics. Replication of fragments up to 200 bp is complete in less than 20s at 60°C. Additionally, the dNTP mix in this FastMix contains dUTP in place of dTTP. Inclusion of uracil-N-glycosylase (UNG) prevents amplification of carry-over contamination from previous dU-containing PCRs.

Instrument Compatibility

Different real-time PCR systems employ different strategies for the normalization of fluorescent signals and correction of well-to-well optical variations. It is critical to match the appropriate qPCR reagent to your specific instrument. PerfeCTa qPCR FastMix, UNG, Low ROX provides seamless integration on the Applied Biosystems 7500, 7500 Fast, ViiA 7, or Stratagene MX series of real-time PCR systems. Please consult the following table, or visit our web site at www.quantabio.com to find the optimal kit for your instrument platform.

Reagent	Cat Nos	Compatible Real-Time PCR Systems
PerfeCTa qPCR FastMix, UNG, ROX	95077-250, 95077-012, 95077-05K	Applied Biosystems 7000, 7300, 7700, 7900, 7900HT, 7900HT Fast, StepOne™, StepOnePlus™
PerfeCTa qPCR FastMix, UNG, Low ROX	95078-250, 95078-012, 95078-05K	Applied Biosystems 7500, 7500 Fast, ViiA™ 7 Stratagene MX4000™, MX3005P™, MX3000P™
PerfeCTa qPCR FastMix, UNG	95076-250, 95076-012, 95076-05K	Bio-Rad CFX96™, CFX384™, iCycler iQ®, iQ™5, MyIQ™ Opticon™, MiniOpticon™, Chromo4™ Cepheid Smart Cycler®; Qiagen/Corbett Rotor-Gene® Eppendorf Mastercycler® ep realplex Roche Applied Science LightCycler® 480

Components

PerfeCTa qPCR FastMix, UNG, Low ROX (2X): 2X reaction buffer containing optimized concentrations of MgCl₂, dNTPs (dATP, dCTP, dGTP, dUTP), AccuFast Taq DNA Polymerase, UNG, ROX Reference Dye and stabilizers.

Storage and Stability

PerfeCTa qPCR FastMix, UNG, Low ROX is stable for 1 year when stored in a constant temperature freezer at -20°C. For convenience, it may be stored unfrozen at +2 to +8°C for up to 6 months.

Repeated freezing and thawing of the FastMix is not recommended. However, the product demonstrated no loss of performance after 20 freeze-thaw cycles or 2 months at +20°C.

Guidelines for qPCR:

- The design of highly specific primers and probes is a critical parameter for successful real-time PCR. The use of computer aided primer design programs is encouraged in order to minimize the potential for internal secondary structure and complementation at 3'-ends within each primer, the primer pair, and primer/probe combinations. For best results, amplicon size should be limited to 65 - 150 bp. Optimal results may require titration of primer concentration between 100 and 900 nM. A final concentration of 300-400 nM each primer and 100-150 nM probe is effective for most applications. However, increasing the concentration of the primer that initiates synthesis of the target strand that is complementary to the probe can improve fluorescent signal for some primer/probe systems. Validate the efficacy and efficiency of any primer/probe set under fast cycling and/or rapid ramp rate protocols before use in qPCR studies.

Guidelines for qPCR continued:

- Preparation of a reaction cocktail is recommended to reduce pipetting errors and maximize assay precision. Assemble the reaction cocktail with all required components except sample template (genomic DNA or cDNA) and dispense equal aliquots into each reaction tube. Add the DNA template to each reaction as the final step. Addition of samples as 5- μ L volumes will improve assay precision.
- Suggested input quantities of template are: cDNA corresponding to 1 pg to 1 μ g of total RNA; 100 pg to 1 μ g genomic DNA
- After sealing each reaction, vortex gently to mix contents. Centrifuge briefly to collect components at the bottom of the reaction tube.

Reaction Assembly

Component	Volume for 20- μ L rxn.	Final Concentration
PerfeCTa qPCR FastMix, UNG, Low ROX (2X)	10 μ L	1x
Forward primer	variable	300 – 400 nM
Reverse primer	variable	300 – 400 nM
Probe	variable	100 – 150 nM
Nuclease-free water	variable	
Template	5 μ L	variable
Final Volume (μ L)	20 μ L	

Reaction Protocol

Incubate complete reaction mix in a real-time thermal detection system as follows:

UNG incubation	45°C, 2 min (optional)
Initial denaturation:	95°C, 30s (see notes below)
PCR cycling (30-45 cycles):	95°C, 1 to 3s 55 – 65°C, 15 to 30 s (collect and analyze data)

Optimal extension time may be amplicon specific. Use 20s at 60°C as a general starting point. The minimum required extension time will vary for different instrument platforms. Consult your instrument instructions.

Full activation of AccuFast Taq DNA polymerase is instantaneous at 95°C; however, optimal initial denaturation time is *template dependent* and will affect qPCR efficiency and sensitivity. Amplification of genomic DNA or supercoiled plasmid DNA targets may require 5 to 10 min at 95°C to fully denature and fragment the template. This minimizes the potential for renaturation of long fragments and/or repetitive sequence regions that can impair PCR. Short double-stranded DNA template (PCR product) or single-stranded DNA template, such as cDNA template generated using the qScript™ cDNA Synthesis Kit or qScript cDNA SuperMix, require as little as 1s at 95°C. Oligo-dT primed cDNA using an RNase H deficient reverse transcriptase may require 2-3 minutes. Use 30s at 95°C as a general starting point.

Quality Control

Kit components are free of contaminating DNase and RNase. PerfeCTa qPCR FastMix, UNG, Low ROX is functionally tested in qPCR. Kinetic analysis must demonstrate linear resolution over six orders of dynamic range ($r^2 > 0.995$) and a PCR efficiency > 90%.

Limited Label Licenses

Use of this product is covered by one or more of the following US patents and corresponding patent claims outside the US: 5,804,375, 5,538,848, 5,723,591, 5,876,930, 6,030,787 and 6,258,569. The purchase of this product includes a limited, non-transferable immunity from suit under the foregoing patent claims for using only this amount of product for the purchaser's own internal research. No right under any other patent claim and no right to perform commercial services of any kind, including without limitation reporting the results of purchaser's activities for a fee or other commercial consideration, is conveyed expressly, by implication, or by estoppel. This product is for research use only. Diagnostic uses under Roche patents require a separate license from Roche. Further information on purchasing licenses may be obtained from the Director of Licensing, Applied Biosystems, 850 Lincoln Centre Drive, Foster City, California 94404, USA.

The purchase of this product includes a limited, non-transferable right to use the purchased amount of the product to perform Applied Biosystems' patented Passive Reference Method for the purchaser's own internal research. No right under any other patent claim and no right to perform commercial services of any kind, including without limitation reporting the results of purchaser's activities for a fee or other commercial consideration, is conveyed expressly, by implication, or by estoppel. This product is for research use only. For information about these rights or on obtaining additional rights, please contact outlicensing@lifetech.com or Out Licensing, Life Technologies, 5791 Van Allen Way, Carlsbad, California 92008.

Licensed to Quanta BioSciences, under U.S. Patent Nos. 5,338,671, 5,587,287, and foreign equivalents for use in research only.

PerfeCTa, FastMix, AccuFast, and AccuStart are trademarks of Quanta BioSciences Inc. TaqMan is a registered trademark of Roche Molecular Systems, Inc. LightCycler is a registered Trademark of Roche. Applied Biosystems, StepOne, StepOnePlus, ViiA, and ROX are trademarks Life Technologies Corporation. Stratagene, MX3000P, MX3005P and MX4000 are trademarks of Agilent Technologies, Inc. Mastercycler is a trademark of Eppendorf. Rotor-Gene is a registered trademark of Qiagen GmbH. SmartCycler is a trademark of Cepheid. CFX96, CFX384, iCycler iQ, iQ5, MyIQ, Opticon, MiniOpticon and Chromo4 are trademarks of Bio-Rad Laboratories.