

Product Information			
sparQ Universal Library Quant Kit			
Part Number	95210-100		
Number of Reactions	100 Reactions		
Reaction Size	20 μL		
Storage Temperature	-25°C to -15°C		
Lot Number	66206366		
Reference Number	053122, 062322, 22032901,		
	22033102, 22033103,		
	22033104, 22033105,		
	22033106, 082119		
Expiration Date	11/30/2023		

Product Description:

The sparQ Universal Library Quant Kit employs qPCR to quantify library molecules that possess the appropriate adapter tag compatible with Illumina® sequencing platforms at each end. This kit is compatible with a variety of real-time PCR instruments, which supports fast cycling protocols with run times under 40 minutes.

84632 sparQ DNA Standard 1 (20 pM), 0.25mL 84633 sparQ DNA Standard 2 (2 pM), 0.25mL 84634 sparQ DNA Standard 3 (0.2 pM), 0.25mL 84635 sparQ DNA Standard 4 (0.02 pM), 0.25mL 84636 sparQ DNA Standard 5 (0.002 pM), 0.25mL 84637 sparQ DNA Standard 6 (0.0002 pM), 0.25mL 84646 Passive ROX Reference Dye (50x), 0.025mL

Compo	nent P	<u>art Nu</u>	mbers:	
84630	Library	Dilutio	n Buffer	(10X).

84630 Library Dilution Buffer (10X), 1.50mL 84645 sparQ Universal Fast Mastermix (1.25X), 8.0mL

Product Specifications		
95210		
Assay	Functional Assay	
Result	Pass	

Quality Control Analysis and Specifications:

qPCR of the DNA Standards are assayed in triplicate reactions with sparQ Fast Mastermix according to the recommended cycling conditions and must generate a standard curve with $R^2 \ge 0.995$ and a PCR efficiency $\ge 90\%$. Dissociation curve analysis must present as a single melt peak with the expected Tm.

Limitations of Use

Quantabio and Ultraplex are registered trademarks of QIAGEN Beverly, Inc. Quanta Biosciences, qScript, Geltrack, ToughMix, PerfeCta, and Fastmix are registered trademarks of Quanta BioSciences Inc. Extracta, AccuStart, AccuMelt, and Accuvue are trademarks of Quanta BioSciences Inc. Applied Biosystems, StepOne, StepOnePlus and ROX are trademarks of Thermo Fisher Scientific and or its subsidiaries. Illumina is a registered trademark of Illumina, Inc. Please contact QIAGEN-Beverly for more information.

This product was developed, manufactured, and sold for *in vitro* use only. The product is not suitable for administration to humans or animals. SDS sheets relevant to this product are available upon request.