

Product Information			
UltraPlex 1-Step ToughMix® Low ROX™			
Part Number	95168-500		
Number of Reactions	500 Reactions		
Reaction Size	20 μL		
Storage Temperature	-25°C to -15°C		
Lot Number	66243906		
Reference Number	111623		
Expiration Date	12/31/2024		

#### **Product Description:**

UltraPlex 1-Step ToughMix Low ROX is a ready-to-use, 4X concentrated master mix 1-step reverse transcription and real-time quantitative PCR (RT-qPCR) of RNA templates using probe-based detection methods. First-strand cDNA synthesis and PCR amplification are carried out in the same tube without opening between steps. Optimized to deliver highly sensitive quantification of RNA viruses or low abundance RNA targets in uni- or highly multiplexed RNA detection assays, this reagent chemistry is optimized to deliver maximum assay sensitivity, precision and reproducibility with miniaturized reaction volumes and either conventional or accelerated thermal cycling protocols. UltraPlex 1-Step ToughMix contains all required components for RTqPCR except RNA template and probe and is compatible with all dual-labeled probe chemistries.

# **Component Part Numbers:**

84381 UltraPlex1-Step ToughMix Low ROX 0.50 mL

Product Specifications				
95168				
Assay	UltraPlex 1-Step ToughMix Functional Assay	DNase	RNase	
			·	

## **Quality Control Analysis and Specifications:**

## **Nuclease Assay:**

**DNase:** DNase activity must be below the detectable limits of 100 pg DNase I equivalent as assayed using a fluorogenic substrate following a 1 hour incubation at 37°C with each kit component at 1X concentration. **RNase:** RNase activity must be below the detectable limits of 1 pg RNase A equivalent as assayed using a fluorogenic substrate following a 1 hour incubation at 37°C with each kit component at 1X concentration.

**UltraPlex 1-Step ToughMix Functional Assay:** Must have detection of  $\beta$  actin, PPIA, GAPDH, and TUBA from 1 pg to 100,000 pg of UHR total RNA. Cq standard curve analysis for each targeted sequence must have a coefficient of determination ( $R^2$ )  $\geq$ 0.990 with a slope between -3.20 and -3.65.

#### Limitations of Use

Quantabio, UltraPlex, qScript, GelTrack, ToughMix, PerfeCTa, and FastMix are registered trademarks of Quantabio, LLC. Applied Biosystems, StepOne, StepOnePlus and ROX are trademarks of Thermo Fisher Scientific and or its subsidiaries. Please contact Quantabio for more information.

This product was developed, manufactured, and sold for *in vitro* use only. The product is not suitable for administration to humans or animals. SDS sheets relevant to this product are available upon request.