

Product Information	
Extracta Plus RNA, 10 rxn	
Part Number	95214-010
Number of Reactions	10 Reactions
Storage Temperature	15°C to 25°C
Lot Number	178029707

Product Description:

The Extracta Plus RNA kits offer rapid purification of Total RNA from animal cells or tissues. The Extracta Plus Spin column workflow enables simultaneous processing of multiple samples in just 25 mins to yield purified total RNA suitable for a range of downstream applications such as next generation sequencing, RT-(q)PCR, cDNA synthesis and microarrays. The kit includes specially designed Extracta Plus DNA removal columns for effective removal of genomic DNA contamination, providing a convenient solution for applications that are sensitive to low amounts of DNA contamination such as RT-qPCR.

Component Part Numbers:

1124037	Buffer EPRL	6 ml
1124040	Buffer EPRW1	7 ml
1124043	Buffer EPRW2	3 ml
1124046	RNase-free Water	10 ml
1124014	Extracta Plus Spin Column	10 pc
1124048	Extracta Plus DNA Removal Column	10 pc
1124050	CollectionTubes 1.5ml	10 pc
1124016	CollectionTubes 2ml	10 pc

Product Specifications	
95214-010	
Assay	Functional Assay
Result	Pass

Quality Control Analysis and Specifications:
Membrane control

The RNA-binding capacity was tested by determining the recovery obtained with 122 - 138 µg of input total carrier RNA on the membrane of the Spin Column. Greater than 80% recovery was obtained.

Pass

As tested by PCR, no PCR inhibitors were released from the membrane.

Pass

Limitations of Use

QuantaBio, UltraPlex, qScript, GelTrack, ToughMix, PerfeCta, and FastMix are registered trademarks of QuantaBio, LLC. Applied Biosystems, StepOne, StepOnePlus and ROX are trademarks of Thermo Fisher Scientific and or its subsidiaries. Please contact QuantaBio, LLC for more information.

This product was developed, manufactured, and sold for *in vitro* use only. The product is not suitable for administration to humans or animals. SDS sheets relevant to this product are available upon request.

100 Cummings Center, Suite 407J, Beverly, MA 01915 • Phone +1 (888) 959 5165 • <https://www.quantabio.com/>

FMWI016.2 Rev 02

Buffers

Conductivity and pH of buffers and solutions were tested and were within the ranges below:

	Conductivity (mS/cm)	pH	
Buffer EPRL	23.95 – 29.07 (1/10 dilution)	6.70 – 7.20	Pass
Buffer EPRW1	29.02 – 33.75 (1/2 dilution)	7.30 – 7.70	Pass
Buffer EPRW2	3.19 – 3.83	7.30 – 7.70	Pass
RNase-free water	≤ 0.005	–	Pass

Functional absence of RNase was checked for RNase-free water and Buffer EPRW2. Pass

Extracta Plus DNA Removal Column

Extracta Plus DNA Removal Columns were functionally tested for DNA binding capacity. Pass

Extracta Plus Spin Column

Extracta Plus Spin Columns were tested by preparation of total RNA from human cells and from 3 mg of tissue.

Average yields and purities obtained were:

Human cells:	≥ 75 µg total RNA	Pass
Kidney:	≥ 6 µg total RNA	Pass
A ₂₆₀ /A ₂₈₀ :	1.9 – 2.3 (at pH 7.5)	Pass

Integrity and size distribution of total RNA was checked by denaturing agarose gel electrophoresis and ethidium bromide staining. Two sharp 28S and 18S ribosomal RNA bands appeared on the stained gel without any smear towards the smaller sized RNAs. Pass

Limitations of Use

QuantaBio, UltraPlex, qScript, GelTrack, ToughMix, PerfeCTa, and FastMix are registered trademarks of QuantaBio, LLC.

Applied Biosystems, StepOne, StepOnePlus and ROX are trademarks of Thermo Fisher Scientific and or its subsidiaries. Please contact QuantaBio, LLC for more information.

This product was developed, manufactured, and sold for *in vitro* use only. The product is not suitable for administration to humans or animals. SDS sheets relevant to this product are available upon request.

100 Cummings Center, Suite 407J, Beverly, MA 01915 • Ph (888) 927-7027 • Fax (978) 867-5724 • <https://www.quantabio.com/> FMWI016.2
Rev 02